

Analytical Method Information

Analyte	MDL	Reporting	Surrogate	Duplicate	Matrix Spike		Blank Spike / LCS	
		Limit	%R	RPD	%R	RPD	%R	RPD
Trihalomethanes DHS 524.2 in Water (EPA 524.2)								
Preservation: Dechlorinate, Store cool at 4°C								
Container: VOA Vial - Na2S2O3								
			Amount Required: 3 VOA vials			Hold Time: 14 days		
Bromodichloromethane	0.14000	1.0000 ug/l			70 - 130	30	70 - 130	30
Bromoform	0.28000	1.0000 ug/l			70 - 130	30	70 - 130	30
Chloroform	0.27000	1.0000 ug/l			70 - 130	30	70 - 130	30
Dibromochloromethane	0.30000	1.0000 ug/l			70 - 130	30	70 - 130	30
Trihalomethanes (total)	0.16000	1.0000 ug/l						
surr: Bromofluorobenzene				70 - 130				
surr: Dibromofluoromethane				70 - 130				
surr: Toluene-d8				70 - 130				