

Analytical Method Information

Analyte	MDL	Reporting Limit	Surrogate %R	Duplicate RPD	Matrix Spike %R	RPD	Blank Spike / LCS %R	RPD
Conductivity-120.1 in Water (EPA 120.1)								
Preservation: Store cool at 4°C								
Container: 500 mL Poly Unpres								
Amount Required: 500 ml								
Hold Time: 28 days								
Specific Conductance (EC)	1.0	20 umhos/cm		10				
Perchlorate Low Level in Water (EPA 314.0)								
Preservation: Store cool at 4°C								
Container: 500 mL Poly Unpres								
Amount Required: 500 ml								
Hold Time: 28 days								
Perchlorate	0.80	2.0 ug/l		20	70 - 130	20	85 - 115	20